FIELD ID NO: _

PART 6. APPLICATION RECORDS (may be used in field and greenhouse trials for calibrations, rechecks, and target checks)

PART 6 PAGE COMPLETE IF APPROPRIATE: "THIS IS A TRUE COPY OF THE ORIGINAL"										Trial Year 2023
ABOV	E DATA ENTI	ERED BY:					DATE:		_	
Are individual nozzle outputs within 5% of the mean during each run?						YES	NO	NA		
If this is a 3-discharge calibration run or a 3-run target check, is each boom discharge rate (far right column in rows 1, 2, and 3) within 5% of the mean?							NO			
•					•	Y ES	NU	NA		
*Indicate whether discharge rate is calculated for: Total Boom Volume If yes, were results within 5% of original calibration or target output?						_ Mean Nozzle Volume YES NO NA				**(A or B)/C=D
έ Τ 1"	4. 1.4. 12.	1	.1. 1.4. 1 (T. 4.1 D.	X7 . 1	Marrin	1. 37.1			**/A D)/C D
Is this a 1-run recheck? ¹ Yes No Is this a 3-run target check? ¹ Yes No						Total	A	В	С	Average Discharge rate** D
3										
2										
1										
	1	2	3	4	3	0	boom volume	outlet	(SCC)	Tate
RUN	-	Nozzle/hopper outlet number along boom (ent diagram fo	r nozzle #s)	Total boom volume	Mean per nozzle or	Time (sec)	Discharge rate*
PRESS	-								<u> </u>	
BRIEF	LY DESCRIBE	E PROCEDUR	E USED TO C	HECK DISCH	HARGE CALII	BRATION				
NSTR	UMENT USEI	O TO MEASU	RE WATER (e	.g. 100 ml gra	duated cylinde	er)				
LOCA	TION WHERE	THE CALIBR	RATION WAS	PERFORMEI)					
EQUIPMENT IDENTIFIER TIME TIME										
EQUIP	MENT IDENT	TIFIER								
	TRUCTIONS: I tions that do no					ify that the ou	tput is within +/	-5% of the mo	st recent full	calibration.
7 77700	EQUIPMENT USED FOR APPLICATION NUMB STRUCTIONS: If conducting a recheck, please provide calculations to verify that the output is within +/-5% of the most reco									